

# Certificate of Conformity

## USP Protocols

Representative samples of Aflex Hose Ltd PTFE hose products have been tested in accordance with USP protocols and are found to conform to the requirements of the following tests:

USP 27, NF22, 2004, for Class VI Plastics at 70°C <88>

USP 27, NF22, 2004, Elution test <87> and are considered non-cytotoxic

The PTFE used in the manufacture of Aflex hose Ltd hose products is manufactured solely from materials, which meet the requirements of USP Plastics Class VI.

Yours faithfully,

For and on behalf of Aflex Hose Limited



Richard C White  
Quality Assurance Manager

6<sup>th</sup> December 2004

**The World's Leading Manufacturer of PTFE Flexible Hose**

Spring Bank Industrial Estate, Watson Mill Lane  
Sowerby Bridge, Halifax. HX6 3BW  
Tel: 01422 317200

Fax Nos: (01422) 836000 – UK Sales/Admin (01422) 833000 – Export Sales  
(01422) 836666 – Purchasing (01422) 839398 – Technical

